



TEST RESULT SUMMARY

EUROPEAN STANDARD EN 60601-1-2: 2001

EUROPEAN TELECOMMUNICATION STANDARD EN 301 489-17, V1.2.1: 2002

MANUFACTURER'S NAME	Neoprobe Corporation
NAME OF EQUIPMENT	Bluetooth Gamma Detection Probes Straight/Angled; Bluetooth Serial Port Adapter
MODEL NUMBER(S) TESTED	Probe – 1100 Probe – 1101 Serial Port Adapter - 1131
MANUFACTURER'S ADDRESS	425 Metro Place North, Suite 300 Dublin OH 43017
TEST REPORT NUMBER	WC605823
TEST DATE(S)	12-13 October 2006

According to testing performed at TÜV Product Service Inc, the above-mentioned unit is in compliance with the electromagnetic compatibility requirements of European Standard EN 60601-1-2: 2001 and European Telecommunication Standard EN 301 489-17, V1.1.1 2002.

It is the manufacturer's responsibility to assure that additional production units of this model are manufactured with identical electrical and mechanical characteristics. Any modifications necessary for compliance made during testing on the above mentioned date(s) must be implemented in all production units for compliance to be maintained.

TÜV Product Service Inc, as an independent testing laboratory, declares that the equipment tested as specified above conforms to the electromagnetic compatibility requirements of European Standard EN 60601-1-2: 2001: "Medical Electrical Equipment, Part 1-2: General Requirements for Safety - Collateral Standard: Electromagnetic Compatibility - Requirements and Tests" and European Telecommunications Standard EN 301 489-17, V1.2.1: 2002: "Electromagnetic Compatibility and Radio Spectrum Matters (ERM); Electromagnetic Compatibility (EMC) standard for Radio Equipment and Services; Part 17: Specific Conditions for 2,4 GHz Wideband Transmission Systems and 5 GHz High Performance TLAN Equipment".

Date: 24 October 2006

Location: New Brighton MN
USA

Greg S Jakubowski
Senior EMC Technician

Joel T Schneider
Senior EMC Engineer

Not Transferable